



A-level PHYSICS (7408/3BB)

Paper 3 – Section B (Medical Physics)

Mark scheme

Mark schemes are prepared by the Lead Assessment Writer and considered, together with the relevant questions, by a panel of subject teachers. This mark scheme includes any amendments made at the standardisation events which all associates participate in and is the scheme which was used by them in this examination. The standardisation process ensures that the mark scheme covers the students' responses to questions and that every associate understands and applies it in the same correct way. As preparation for standardisation each associate analyses a number of students' scripts. Alternative answers not already covered by the mark scheme are discussed and legislated for. If, after the standardisation process, associates encounter unusual answers which have not been raised they are required to refer these to the Lead Assessment Writer.

It must be stressed that a mark scheme is a working document, in many cases further developed and expanded on the basis of students' reactions to a particular paper. Assumptions about future mark schemes on the basis of one year's document should be avoided; whilst the guiding principles of assessment remain constant, details will change, depending on the content of a particular examination paper.

Further copies of this mark scheme are available from aqa.org.uk

Physics – Mark scheme instructions to examiners

1. General

The mark scheme for each question shows:

- the marks available for each part of the question
- the total marks available for the question
- the typical answer or answers which are expected
- extra information to help the Examiner make his or her judgement and help to delineate what is acceptable or not worthy of credit or, in discursive answers, to give an overview of the area in which a mark or marks may be awarded.

The extra information is aligned to the appropriate answer in the left-hand part of the mark scheme and should only be applied to that item in the mark scheme.

At the beginning of a part of a question a reminder may be given, for example: where consequential marking needs to be considered in a calculation; or the answer may be on the diagram or at a different place on the script.

In general the right-hand side of the mark scheme is there to provide those extra details which confuse the main part of the mark scheme yet may be helpful in ensuring that marking is straightforward and consistent.

2. Emboldening

- 2.1** In a list of acceptable answers where more than one mark is available ‘any **two** from’ is used, with the number of marks emboldened. Each of the following bullet points is a potential mark.
- 2.2** A bold **and** is used to indicate that both parts of the answer are required to award the mark.
- 2.3** Alternative answers acceptable for a mark are indicated by the use of **or**. Different terms in the mark scheme are shown by a / ; eg **allow** smooth / free movement.

3. Marking points

3.1 Marking of lists

This applies to questions requiring a set number of responses, but for which candidates have provided extra responses. The general principle to be followed in such a situation is that ‘right + wrong = wrong’.

Each error / contradiction negates each correct response. So, if the number of errors / contradictions equals or exceeds the number of marks available for the question, no marks can be awarded.

However, responses considered to be neutral (often prefaced by ‘ignore’ in the mark scheme) are not penalised.

3.2 Marking procedure for calculations

Full marks can usually be given for a correct numerical answer without working shown unless the question states 'Show your working'. However, if a correct numerical answer can be evaluated from incorrect physics then working will be required. The mark scheme will indicate both this and the credit (if any) that can be allowed for the incorrect approach.

However, if the answer is incorrect, mark(s) can usually be gained by correct substitution / working and this is shown in the 'extra information' column or by each stage of a longer calculation.

A calculation must be followed through to answer in decimal form. An answer in surd form is never acceptable for the final (evaluation) mark in a calculation and will therefore generally be denied one mark.

3.3 Interpretation of 'it'

Answers using the word 'it' should be given credit only if it is clear that the 'it' refers to the correct subject.

3.4 Errors carried forward, consequential marking and arithmetic errors

Allowances for errors carried forward are likely to be restricted to calculation questions and should be shown by the abbreviation ECF or *conseq* in the marking scheme.

An arithmetic error should be penalised for one mark only unless otherwise amplified in the marking scheme. Arithmetic errors may arise from a slip in a calculation or from an incorrect transfer of a numerical value from data given in a question.

3.5 Phonetic spelling

The phonetic spelling of correct scientific terminology should be credited (eg fizix) **unless** there is a possible confusion (eg defraction/refraction) with another technical term.

3.6 Brackets

(.....) are used to indicate information which is not essential for the mark to be awarded but is included to help the examiner identify the sense of the answer required.

3.7 Ignore / Insufficient / Do not allow

'Ignore' or 'insufficient' is used when the information given is irrelevant to the question or not enough to gain the marking point. Any further correct amplification could gain the marking point.

'Do **not** allow' means that this is a wrong answer which, even if the correct answer is given, will still mean that the mark is not awarded.

3.8 Significant figure penalties

An A-level paper may contain up to 2 marks (1 mark for AS) that are contingent on the candidate quoting the **final** answer in a calculation to a specified number of significant figures (sf). This will generally be assessed to be the number of sf of the datum with the least number of sf from which the answer is determined. The mark scheme will give the range of sf that are acceptable but this will normally be the sf of the datum (or this sf -1).

3.9 Unit penalties

An A-level paper may contain up to 2 marks (1 mark for AS) that are contingent on the candidate quoting the correct unit for the answer to a calculation. The need for a unit to be quoted will be indicated in the question by the use of 'State an appropriate SI unit for

your answer '. Unit answers will be expected to appear in the most commonly agreed form for the calculation concerned; strings of fundamental (base) units would not. For example, 1 tesla and 1 weber/metre² would both be acceptable units for magnetic flux density but 1 kg m² s⁻² A⁻¹ would not.

3.10 Level of response marking instructions.

Level of response mark schemes are broken down into three levels, each of which has a descriptor. The descriptor for the level shows the average performance for the level. There are two marks in each level.

Before you apply the mark scheme to a student's answer read through the answer and annotate it (as instructed) to show the qualities that are being looked for. You can then apply the mark scheme.

Determining a level

Start at the lowest level of the mark scheme and use it as a ladder to see whether the answer meets the descriptor for that level. The descriptor for the level indicates the different qualities that might be seen in the student's answer for that level. If it meets the lowest level then go to the next one and decide if it meets this level, and so on, until you have a match between the level descriptor and the answer. With practice and familiarity you will find that for better answers you will be able to quickly skip through the lower levels of the mark scheme.

When assigning a level you should look at the overall quality of the answer and not look to pick holes in small and specific parts of the answer where the student has not performed quite as well as the rest. If the answer covers different aspects of different levels of the mark scheme you should use a best fit approach for defining the level and then use the variability of the response to help decide the mark within the level, i.e. if the response is predominantly level 2 with a small amount of level 3 material it would be placed in level 2.

The exemplar materials used during standardisation will help you to determine the appropriate level. There will be an answer in the standardising materials which will correspond with each level of the mark scheme. This answer will have been awarded a mark by the Lead Examiner. You can compare the student's answer with the example to determine if it is the same standard, better or worse than the example. You can then use this to allocate a mark for the answer based on the Lead Examiner's mark on the example.

You may well need to read back through the answer as you apply the mark scheme to clarify points and assure yourself that the level and the mark are appropriate.

Indicative content in the mark scheme is provided as a guide for examiners. It is not intended to be exhaustive and you must credit other valid points. Students do not have to cover all of the points mentioned in the indicative content to reach the highest level of the mark scheme

An answer which contains nothing of relevance to the question must be awarded no marks.

Question	Answers	Additional Comments/Guidance	Mark
01.1	Principal focus is the point on the principal axis through which rays which were parallel to the principal axis pass after refraction by the lens ✓ Power is reciprocal of focal length measured in m ✓	Allow suitable labelled diagram Allow 1/f measured in m	1 1
01.2	First correct ray ✓ Second correct ray with labelled image ✓		1 1
01.3	Myopia or short sight ✓		1
01.4	$1/-0.33 = 1/0.25 + 1/v$ ✓ $v = (-)0.14 \text{ m}$ ✓		1 1
01.5	Cones active/simulated ✓ Cones stimulated by images must be separated by at least 1 unstimulated cone ✓		1 1

02.1	Frequency does not change ✓ Amplitude is reduced ✓		1 1
02.2	Ossicles lever system produces increase in force ✓ Area of oval window much less than area of ear drum ✓ Pressure = F/A so large increase in pressure ✓		1 1 1
02.3	$I = 1.0 \times 10^{-12} \text{ W m}^{-2}$ ✓ $I = 1.6 \times 10^{-4} \text{ W m}^{-2}$ ✓	$\times 10^{8.2}$ ✓	1 1 1
02.4	$P = 1.6 \times 10^{-4} \times 4 \times \pi \times 2.0^2$ ✓ $P = 8.0 \times 10^{-3} \text{ W}$ ✓		1 1

3	<p>The mark scheme gives some guidance as to what statements are expected to be seen in a 1 or 2 mark (L1), 3 or 4 mark (L2) and 5 or 6 mark (L3) answer. Guidance provided in section 3.10 of the 'Mark Scheme Instructions' document should be used to assist in marking this question.</p>		<p>The following comparisons are likely to be present:</p> <table border="1" data-bbox="279 280 1412 1041"> <tr> <td data-bbox="279 884 422 1041">Patient Safety</td> <td data-bbox="279 705 422 884">Ionising radiation exposure</td> <td data-bbox="279 492 422 705">PET</td> <td data-bbox="279 280 422 492">u/s</td> </tr> <tr> <td data-bbox="422 884 598 1041">Convenience</td> <td data-bbox="422 705 598 884">Scan time equipment cost</td> <td data-bbox="422 492 598 705">2-4 h Large, bulky expensive</td> <td data-bbox="422 280 598 492">None (no tracer) 10 -15 min portable relatively cheap</td> </tr> <tr> <td data-bbox="598 884 670 1041">Pre-treatment</td> <td data-bbox="598 705 670 884">Patient movement</td> <td data-bbox="598 492 670 705">Needs tracer injection Must lie still</td> <td data-bbox="598 280 670 492">No injection Movement tolerated</td> </tr> <tr> <td data-bbox="670 884 742 1041">Discomfort</td> <td data-bbox="670 705 742 884">Patient movement</td> <td data-bbox="670 492 742 705">Not good if claustrophobic</td> <td data-bbox="670 280 742 492">Requires cold gel</td> </tr> <tr> <td data-bbox="742 884 845 1041">Information</td> <td data-bbox="742 705 845 1041"></td> <td data-bbox="742 492 845 1041">Chemical and physiological changes related to metabolism Useful for detecting brain activity</td> <td data-bbox="742 280 845 1041">Size position movement (of foetus, organs etc.) Cannot penetrate bone – cannot examine brain Good for imaging soft tissues Cannot pass</td> </tr> </table>	Patient Safety	Ionising radiation exposure	PET	u/s	Convenience	Scan time equipment cost	2-4 h Large, bulky expensive	None (no tracer) 10 -15 min portable relatively cheap	Pre-treatment	Patient movement	Needs tracer injection Must lie still	No injection Movement tolerated	Discomfort	Patient movement	Not good if claustrophobic	Requires cold gel	Information		Chemical and physiological changes related to metabolism Useful for detecting brain activity	Size position movement (of foetus, organs etc.) Cannot penetrate bone – cannot examine brain Good for imaging soft tissues Cannot pass
Patient Safety	Ionising radiation exposure	PET		u/s																			
Convenience	Scan time equipment cost	2-4 h Large, bulky expensive		None (no tracer) 10 -15 min portable relatively cheap																			
Pre-treatment	Patient movement	Needs tracer injection Must lie still		No injection Movement tolerated																			
Discomfort	Patient movement	Not good if claustrophobic		Requires cold gel																			
Information		Chemical and physiological changes related to metabolism Useful for detecting brain activity		Size position movement (of foetus, organs etc.) Cannot penetrate bone – cannot examine brain Good for imaging soft tissues Cannot pass																			
Mark	Criteria	QoWC																					
6	<p>All three aspects covered:</p> <p>An comparison of patient safety in terms of ionising radiation exposure.</p> <p>An appreciation of convenience in terms of pre treatment, scan time, patient movement and discomfort. There may also be a discussion of comparative cost and equipment partability.</p> <p>A discussion of the types of information available in terms of uses and limitations.</p>	<p>The student presents relevant information coherently, employing structure, style and sp&g to render meaning clear. The text is legible.</p>																					
5	<p>Two of the three aspects fully covered, with some detail missing from the third.</p>																						
4	<p>One aspect fully covered, with some detail missing from the other two</p> <p>Or</p>	<p>The student presents relevant information and in a way which assists the communication of meaning. The text is</p>																					

	<p>Two aspects fully covered, with little or no relevant information about the third.</p> <p>3</p> <p>All three aspects partially covered, with some detail missing from each</p> <p>Or</p> <p>One aspect fully covered, with little or no relevant information about the other two</p> <p>2</p> <p>Two aspects partially covered, with little or no relevant information about the third.</p> <p>1</p> <p>One aspect partially covered, with little or no relevant information about the other two.</p> <p>0</p> <p>Little or no relevant information about any of the three aspects.</p>	<p>legible. Sp&g are sufficiently accurate not to obscure meaning.</p> <p>The student presents some relevant information in a simple form. The text is usually legible. Sp&g allow meaning to be derived although errors are sometimes obstructive.</p> <p>The student's presentation, spelling punctuation and grammar seriously obstruct understanding.</p>	<table border="1"> <tr> <td data-bbox="201 884 1085 1052"></td> <td data-bbox="201 705 1085 884"></td> <td data-bbox="201 492 1085 705">Can provide info re malignancy and tumour spreading</td> <td data-bbox="201 268 1085 492">through air spaces/lungs Cannot distinguish between benign and malignant solid masses</td> </tr> </table>			Can provide info re malignancy and tumour spreading	through air spaces/lungs Cannot distinguish between benign and malignant solid masses	
		Can provide info re malignancy and tumour spreading	through air spaces/lungs Cannot distinguish between benign and malignant solid masses					

04.1	<p>The physical half-life depends only on the properties of the radioactive nuclide ✓ biological removal of the nuclide also occurs (thus removing the nuclide more quickly overall) WTTE ✓</p>		1 1
04.2	<p>$A_t = A_0 e^{-\lambda t}$ correct sub 1200 = 2700 e^{-5λ} ✓ $(\lambda_E = \ln(2700 / 1200) / 5 = 0.1622)$ $T_E = \ln(2) / 0.1622 = 4.273$ ✓ $1/4.273 = 1/20 + 1/T_B$ ✓ $T_B = 5.4$ days ✓</p>		1 1 1 1
04.3	<p>Beta more strongly ionising than gamma so ¹³¹I more likely to damage cells/increase radiation dose ✓ Gamma rays for ¹³¹I are over 4 x more energetic which can cause problems when imaging with a gamma camera ✓ 190 h >> 6.0 h so with ¹³¹I body will remain radioactive for longer posing a greater danger to patient and others he/she in contact with ✓ Half-life of ⁹⁹T^m may be too short for certain types of diagnosis to be undertaken ✓ Sensible conclusion based on above points ✓ 4 marks max</p>	<p>Conclusion may refer to radionuclide ending up at right place in body eg ¹³¹I taken up by thyroid, ⁹⁹T^m taken up by bone and red blood cells</p>	1 1 1 1

